



Infection Prevention with Confidence!



In the war against microbes, Accel® Disinfectants are the first line of defense for Infection Prevention. Accel® provides comfort and confidence that commonly touched surfaces and instruments are rendered safe to use.

PREvention is designed for use in instrument and device reprocessing of semi-critical and critical devices where a high level disinfection and rapid contact time are required. PREvention carries a 5 minute high level disinfection claim against bacteria, viruses, fungi, mycobacteria and spores.

EFFICACY OF PREVENTION

SPORICIDAL (6 hours at 20°C): *Bacillus subtilis*, *Clostridium sporogenes*

TUBERCULOCIDAL (5 min at 20°C): *Mycobacterium terrae*

FUNGICIDAL (5 min at 20°C): *Trychophyton mentagrophytes*

BACTERICIDAL (5 min) at 20°C: *Pseudomonas aeruginosa*, *Staphylococcus aureus*, *Salmonella choleraesuis*

VIRUCIDAL (5 min at 20°C): Poliovirus, Sabin strain type 1, which allows for a General Virucide claim against most enveloped and non-enveloped viruses.

DIRECTIONS FOR USE

(HEAVILY SOILED SURFACES REQUIRE CLEANING PRIOR TO DISINFECTION)

CRITICAL ITEMS: Chemosterilization for instruments and devices that come in contact with sterile tissue, the blood stream & recirculating body fluids – such as surgical instruments and foot care instruments. Place pre-cleaned, rinsed and dried instruments into a container of undiluted PREvention. Allow the instruments to soak for at least 6 hours. Remove and rinse with sterile water. (See Notice To User)

SEMI-CRITICAL ITEMS: High-Level Disinfection for instruments and devices that come in contact with non-intact skin or mucous membranes that are not normally considered sterile such as respiratory equipment, surgical mirrors etc. Place pre-cleaned, rinsed and dried instruments into a container of undiluted PREvention. Allow the instruments to soak for at least 5 minutes. Remove and rinse with sterile or potable water. (See Notice To User)

14 DAY RE-USE CLAIM: This solution may be used and re-used for up to 14 days. Solution in its original packaging is good until date of expiry. Test Strips (AHP2) for confirming the active ingredient and the stability of PREvention are available.

NOTICE TO USER: Please ensure the instruments, equipment or device material to be disinfected is compatible with PREvention. Compatible materials include: acrylic, silicone rubber, low & high density plastics and rigid stainless steel instruments (visit the Compatibility Section on www.virox.com for a full list of compatible materials). Continuous exposure to copper and brass should be avoided. PREvention is not recommended for use on carbon steel and anodized aluminum. If you are uncertain as to the material composition of the device or instrument, confirm with the manufacturer before proceeding.

Accel PREvention

Product Number	ACCHLD5P-4
DIN #	02279274
Concentration of Active	2% Hydrogen Peroxide
Container Size	4L
Dilution	Ready to Use
Shelf Life	12 Months



Engineering Revolutionary Disinfectants for the War Against Microbes

1 (800) 387-7578
www.virox.com