



SciCan
1440 Don Mills Road
Toronto, Ontario
Canada
M3B 3P9

2008-01-11
Prof. We/ku

OPTIM[®] 33TB
DIN EN 14348 (April 2005)
Mycobactericidal activity
phase 2, step 1

TEST REPORT

Identification of the test laboratory: SN 7394

Test product: OPTIM[®] 33TB

Batch number: Lot. # 7379

Manufacturer: SciCan

Date of order: 2007-12-04

Date of delivery: 2007-12-10

Storage conditions: those of the manufacturer

Appearance: clear colourless liquid

Odour: aromatic

Active substance: not indicated

pH-values (DIN 19266): 96%: 2.72

SN 7394 DIN EN 14348 page 1 of 7

test method:	DIN EN 14348 (April 2005) Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants in the medical area including instrument disinfectants (phase 2, step 1)
period of analysis:	2007-12-20 to 2008-01-10
test temperature:	20 °C ± 1 °C
product test concentrations:	96 % (v/v end concentration) Note: The product was prepared with a higher concentration of 120%, therefore the end concentration after dilution during the test (interfering substance and test suspension) was 96% (v/v).
contact times:	1, 5 and 10 minutes
interfering substances:	0.3% albumine + 0.3% sheep erythrocytes (dirty conditions)
inactivation method:	dilution-neutralisation
neutralizer:	1.0 % polysorbate 80 + 3.0 % saponine + 0.3 % lezithine
counting procedure:	spread plate method
diluent used for product test solution:	distilled water
stability and appearance of the mixture during the procedure:	no precipitate or flocculent
incubation:	36 °C ± 1 °C – 21 d
test strains:	<i>Mycobacterium terrae</i> ATCC 15755 <i>Mycobacterium avium</i> ATCC 15769

**Test results
with OPTIM® 33TB
according to DIN EN 14348**

test strain: *Mycobacterium terrae* ATCC 15755

interfering substance: **0.3% albumine + 0.3% sheep erythrocytes
(dirty conditions)**

validation

Validation- suspension N _v and N _{v0}	Contact time (min.)	Validation		
		Control (A)	Control (B)	Control (C)
10 ⁻¹ Vc1: 49+58 10 ⁻¹ Vc2: 62+45 N _{v0} : 107	1	Vc1: 67+43 Vc2: 70+46 A: 113	Vc1: 40+64 Vc2: 61+56 B: 110.5	Vc1: 62+50 Vc2: 51+54 C: 108.5
10 ⁻¹ Vc1: 49+58 10 ⁻¹ Vc2: 62+45 N _{v0} : 107	5	Vc1: 52+56 Vc2: 60+56 A: 112	Vc1: 40+64 Vc2: 61+56 B: 110.5	Vc1: 48+68 Vc2: 42+58 C: 108
10 ⁻¹ Vc1: 49+58 10 ⁻¹ Vc2: 62+45 N _{v0} : 107	10	Vc1: 52+66 Vc2: 48+47 A: 106.5	Vc1: 40+64 Vc2: 61+56 B: 110.5	Vc1: 48+48 Vc2: 49+54 C: 99.5

Verification:

N_{v0} is between 30 and 160 cfu/ml (3.0x10¹ and 1.6x10²)

A, B, C is equal to or greater than 0,5 x N_{v0},

Legend:

N_{v0} is number of cfu/ml of the test suspension in the validation test mixture

A is number of cfu/ml of the experimental condition control

B is number of cfu/ml of the neutralizer non-toxicity control

C is number of cfu/ml of the dilution neutralization test control

**Test results
with OPTIM® 33TB
according to DIN EN 14348**

test strain: *Mycobacterium terrae* ATCC 15755

interfering substance: 0.3% albumine + 0.3% sheep erythrocytes
(dirty conditions)

test suspension and main test

test suspension N		
	Vc1	Vc2
10 ⁻⁷	260+241	201+244
10 ⁻⁸	34+31	26+39
N:	4.89x10 ⁹	
N ₀ :	4.89x10 ⁸	
lg N ₀ :	8.69	

Test proce-dure at concen- tration (v/v)		after contact time (min)					
		1		5		10	
				Vc1	Vc2	Vc1	Vc2
96%	10 ⁰ :	>330+>330	>330+>330	0+0	0+0	0+0	0+0
	10 ⁻¹ :	>330+>330	>330+>330	0+0	0+0	1+0	0+0
	10 ⁻² :	>330+>330	>330+>330	0+0	0+0	0+0	0+0
	10 ⁻³ :	96+41	28+120	0+0	0+0	0+0	0+0
	N _a :	1.42x10 ⁶		<140		<140	
	lgN _a :	6.15		<2.15		<2.15	
	lgR:	2.54		>6.54		>6.54	

Verification:

N is between 1.5x10⁹ and 5.0 x 10⁹ cfu/ml (9.17 ≤ lgN ≤ 9.70)

N₀ is between 1.5x10⁸ cfu/ml and 5.0x10⁸ cfu/ml (8.17 ≤ lg N₀ ≤ 8.70)

Legend:

N is number of cfu/ml of the test suspension

N₀ is number of cfu/ml of the test suspension in the test mixture

R is reduction of viability

N_a is number of cfu/ml of the test mixture

**Test results
with OPTIM® 33TB
according to DIN EN 14348**

test strain: *Mycobacterium avium* ATCC 15769

interfering substance: **0.3% albumine + 0.3% sheep erythrocytes
(dirty conditions)**

validation

Validation- suspension N _v and N _{v0}	Contact time (min.)	Validation		
		Control (A)	Control (B)	Control (C)
10 ⁻¹ Vc1: 38+50 10 ⁻¹ Vc2: 36+66 N _{v0} : 95	1	Vc1: 88+80 Vc2: 78+68 A: 157	Vc1: 48+76 Vc2: 51+60 B: 117.5	Vc1: 49+64 Vc2: 38+50 C: 100.5
10 ⁻¹ Vc1: 38+50 10 ⁻¹ Vc2: 36+66 N _{v0} : 95	5	Vc1: 74+40 Vc2: 68+48 A: 115	Vc1: 48+76 Vc2: 51+60 B: 117.5	Vc1: 50+74 Vc2: 46+62 C: 116
10 ⁻¹ Vc1: 38+50 10 ⁻¹ Vc2: 36+66 N _{v0} : 95	10	Vc1: 60+70 Vc2: 59+60 A: 124.5	Vc1: 48+76 Vc2: 51+60 B: 117.5	Vc1: 48+46 Vc2: 44+37 C: 87.5

Verification:

N_{v0} is between 30 and 160 cfu/ml (3.0x10¹ and 1.6x10²)
A, B, C is equal to or greater than 0,5 x N_{v0}.

Legend:

N_{v0} is number of cfu/ml of the test suspension in the validation test mixture
A is number of cfu/ml of the experimental condition control
B is number of cfu/ml of the neutralizer non-toxicity control
C is number of cfu/ml of the dilution neutralization test control

**Test results
with OPTIM® 33TB
according to DIN EN 14348**

test strain: *Mycobacterium avium* ATCC 15769

interfering substance: 0.3% albumine + 0.3% sheep erythrocytes
(dirty conditions)

test suspension and main test

test suspension N		
	Vc1	Vc2
10 ⁻⁷	160+158	201+202
10 ⁻⁸	18+46	27+38
N:	3,91x10 ⁹	
N ₀ :	3,91x10 ⁸	
lg N ₀ :	8.59	

Test proce-dure at concen- tration (v/v)		after contact time (min)					
		1		5		10	
				Vc1	Vc2	Vc1	Vc2
96%	10 ⁰ :	0+0	0+0	0+0	0+0	0+0	0+0
	10 ⁻¹ :	0+0	0+0	0+0	0+0	0+0	0+0
	10 ⁻² :	0+0	0+0	0+0	0+0	0+0	0+0
	10 ⁻³ :	0+0	0+0	0+0	0+0	0+0	0+0
	N _a :	<140		<140		<140	
	lgN _a :	<2.15		<2.15		<2.15	
	lgR:	>6.44		>6.44		>6.44	

Verification:

N is between 1.5x10⁹ and 5.0 x 10⁹ cfu/ml (9.17 ≤ lgN ≤ 9.70)

N₀ is between 1.5x10⁸ cfu/ml and 5.0x10⁸ cfu/ml (8.17 ≤ lg N₀ ≤ 8.70)

Legend:

N is number of cfu/ml of the test suspension

N₀ is number of cfu/ml of the test suspension in the test mixture

R is reduction of viability

N_a is number of cfu/ml of the test mixture

Conclusion:

According to DIN EN 14348 (April 2005), the batch Lot. # 7379 of product OPTIM® 33TB possesses a mycobactericidal activity under dirty conditions at 20°C when diluted at 96% (v/v) for the referenced test strain *Mycobacterium terrae* in 5 minutes and for the referenced test strain *Mycobacterium avium* in 1 minute.

Archive:

The raw data with respect to this test and a copy of the report will be maintained by HygCen in the archive.

Information:

The test results are valid for the named test subject only. Reproduction of any part of this report requires the written permission HygCen GmbH.



Prof. Dr. med. H.-P. Werner
Manager of scientific-technical affairs



K. Naujox
Department manager